
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001876
8:15 22/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:58
STERILIZING 3:07
136.4 C 224kPa 3:07
136.2 C 223kPa 3:37
136.4 C 225kPa 4:07
136.4 C 225kPa 4:37
136.1 C 222kPa 5:07
136.4 C 225kPa 5:37
136.4 C 223kPa 6:07
136.3 C 224kPa 6:37
136.3 C 224kPa 6:37
VENTING 6:37
AIR DRYING 7:28
CYCLE COMPLETE 7:39

Digital Signature #
38BAC667A10F47E2
