
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005597
16:13 24/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:20
STERILIZING 3:24
136.4 C 220kPa 3:24
136.2 C 220kPa 3:54
136.2 C 222kPa 4:24
136.4 C 222kPa 4:54
136.2 C 221kPa 5:24
136.5 C 224kPa 5:54
136.4 C 225kPa 6:24
136.4 C 222kPa 6:54
136.4 C 222kPa 6:54
VENTING 6:55
AIR DRYING 7:49
CYCLE COMPLETE 11:20

Digital Signature #
789B2C264D1E9E55
