
Midmark M11 - v1.0.0
Total Cycles: 4110

Name: RITTER 1
Sterilizer ID: V2178198
12/12/2023 03:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.1 F 0.0
2:00 91.4 F 0.5
4:00 96.2 F 2.1
6:00 122.0 F 4.4
8:00 152.9 F 7.0
10:00 194.5 F 10.3
12:00 199.0 F 8.9
14:00 212.9 F 5.7
16:00 223.7 F 9.2
18:00 235.3 F 13.1
20:00 232.0 F 8.7
22:00 242.4 F 13.4
24:00 240.7 F 11.2
26:00 252.2 F 16.8
28:00 262.3 F 22.7
30:00 271.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.6 F 29.5
2:00 271.6 F 28.9
3:00 271.5 F 29.3
4:00 273.3 F 29.8

Min 271.0 F 28.7
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 226.5 F 0.0
10:00 225.2 F 0.0
15:00 233.1 F 0.0
20:00 226.0 F 0.0
25:00 235.9 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 30:34
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:09:20
CYCLE COMPLETE

12/12/2023 04:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____