
Midmark M11 - v1.0.0
Total Cycles: 2801

Name: RITTER 1
Sterilizer ID: V2178198
01/20/2023 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.9 F 0.0
2:00 82.1 F 0.0
4:00 88.0 F 0.0
6:00 117.4 F 0.0
8:00 149.0 F 1.2
10:00 177.2 F 4.4
12:00 196.4 F 7.7
14:00 210.7 F 11.5
16:00 218.0 F 4.7
18:00 229.8 F 9.1
20:00 239.2 F 13.4
22:00 238.0 F 10.3
24:00 236.0 F 8.2
26:00 248.4 F 14.5
28:00 259.8 F 20.8
30:00 270.3 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 29.0
2:00 271.5 F 28.5
3:00 272.4 F 29.0
4:00 271.5 F 28.4

Min 271.4 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 209.8 F 0.0
10:00 216.5 F 0.0
15:00 218.2 F 0.0
20:00 222.3 F 0.0
25:00 227.5 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:51
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:09:04
CYCLE COMPLETE

01/20/2023 12:38 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____