
Midmark M11 - v1.0.0
Total Cycles: 1183

Name: RITTER 1
Sterilizer ID: V2178198
09/16/2021 02:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.2 F 0.0
2:00 155.2 F 2.4
4:00 182.1 F 6.1
6:00 201.7 F 11.2
8:00 221.6 F 5.9
10:00 235.5 F 11.0
12:00 231.6 F 7.4
14:00 245.7 F 13.5
16:00 243.7 F 11.9
18:00 256.5 F 18.6
20:00 268.0 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.7 F 28.7
2:00 271.5 F 28.5
3:00 272.4 F 28.7
4:00 271.5 F 28.5

Min 271.2 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 233.2 F 0.0
10:00 213.2 F 0.0
15:00 198.2 F 0.0
20:00 199.4 F 0.0
25:00 204.9 F 0.0
30:00 186.8 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:18
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:59:36
CYCLE COMPLETE

09/16/2021 03:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____