
Midmark M11 - v1.0.0
Total Cycles: 2564

Name: RITTER 1
Sterilizer ID: V2178198
11/16/2022 08:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.3 F 0.0
2:00 79.0 F 0.0
4:00 94.5 F 0.0
6:00 119.7 F 0.0
8:00 151.3 F 0.9
10:00 176.2 F 4.4
12:00 193.3 F 8.0
14:00 210.7 F 11.6
16:00 217.9 F 4.7
18:00 227.1 F 9.1
20:00 238.3 F 13.4
22:00 238.0 F 10.2
24:00 235.9 F 8.3
26:00 248.0 F 14.3
28:00 259.3 F 20.5
30:00 269.8 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 29.0
2:00 271.6 F 28.2
3:00 272.4 F 29.1
4:00 272.3 F 28.7

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.1 F 0.0
5:00 213.2 F 0.0
10:00 212.1 F 0.0
15:00 211.2 F 0.0
20:00 198.2 F 0.0
25:00 214.5 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:00
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:08:58
CYCLE COMPLETE

11/16/2022 10:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____