
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000957
10:00 02/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:51
STERILIZING 5:27
136.4 C 232kPa 5:27
136.4 C 230kPa 5:57
136.3 C 231kPa 6:27
136.4 C 233kPa 6:57
136.3 C 231kPa 7:27
136.4 C 230kPa 7:57
136.3 C 231kPa 8:27
136.4 C 232kPa 8:57
136.4 C 232kPa 8:57
VENTING 8:57
AIR DRYING 9:53
CYCLE COMPLETE 24:53

Digital Signature #
D10B076F33184A8D
