
Midmark M11 - v1.0.0
Total Cycles: 3154

Name: RITTER 1
Sterilizer ID: V2178198
04/24/2023 02:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.8 F 0.0
2:00 181.8 F 0.0
4:00 205.5 F 1.7
6:00 219.4 F 6.9
8:00 233.1 F 12.0
10:00 234.8 F 9.0
12:00 238.6 F 9.7
14:00 244.0 F 11.8
16:00 247.5 F 13.8
18:00 260.8 F 21.1
20:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.5 F 28.6
2:00 271.4 F 27.9
3:00 272.6 F 28.8
4:00 271.8 F 28.2

Min 271.3 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 216.4 F 0.0
10:00 240.8 F 0.0
15:00 237.7 F 0.0
20:00 225.5 F 0.0
25:00 222.9 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 20:24
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

04/24/2023 03:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____