

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000953  
7:17 10/12/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:38  
STERILIZING 3:42  
136.0 C 220kPa 3:42  
136.4 C 224kPa 4:12  
136.3 C 221kPa 4:42  
136.1 C 218kPa 5:12  
136.3 C 221kPa 5:42  
136.5 C 224kPa 6:12  
136.4 C 222kPa 6:42  
136.4 C 221kPa 7:12  
136.4 C 221kPa 7:12  
VENTING 7:12  
AIR DRYING 8:03  
CYCLE COMPLETE 8:16

-----  
Digital Signature #  
358EB63780EA2E10  
-----