
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004053
11:24 19/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:47
STERILIZING 3:52
136.0 C 220kPa 3:52
136.4 C 224kPa 4:22
136.2 C 223kPa 4:52
136.2 C 222kPa 5:22
136.1 C 220kPa 5:52
136.3 C 224kPa 6:22
136.3 C 224kPa 6:52
136.3 C 222kPa 7:22
136.3 C 222kPa 7:22
VENTING 7:22
AIR DRYING 8:21
CYCLE COMPLETE 8:31

Digital Signature #
5BF1C004B373B3B2
