
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001964
10:29 16/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:58
STERILIZING 3:02
136.4 C 225kPa 3:02
136.4 C 226kPa 3:32
136.4 C 225kPa 4:02
136.4 C 226kPa 4:32
136.3 C 225kPa 5:02
136.2 C 223kPa 5:32
136.4 C 226kPa 6:02
136.3 C 225kPa 6:32
136.3 C 225kPa 6:32
VENTING 6:32
AIR DRYING 7:23
CYCLE COMPLETE 16:16

Digital Signature #
38C3B2A1B88EA18B
