

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1582

Name: RITTER2  
Sterilizer ID: V2313228  
10/12/2022 08:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.3 F 0.0  
2:00 164.9 F 2.4  
4:00 169.7 F 5.4  
6:00 192.5 F 10.0  
8:00 213.6 F 6.9  
10:00 226.1 F 9.4  
12:00 232.4 F 9.6  
14:00 239.3 F 11.9  
16:00 238.7 F 10.5  
18:00 251.9 F 17.0  
20:00 263.4 F 23.8  
22:00 272.0 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.4 F 29.9  
2:00 271.5 F 29.1  
3:00 272.3 F 30.1  
4:00 272.0 F 29.6

Min 271.2 F 29.1  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.4 F 0.0  
5:00 218.0 F 0.0  
10:00 236.0 F 0.0  
15:00 227.8 F 0.0  
20:00 220.7 F 0.0  
25:00 231.9 F 0.0  
30:00 225.4 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 22:09  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:00:21  
CYCLE COMPLETE

10/12/2022 09:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_