

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2379

Name: RITTER 1  
Sterilizer ID: V2178198  
09/27/2022 01:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.4 F 0.0  
2:00 196.4 F 4.3  
4:00 209.2 F 9.6  
6:00 222.7 F 6.3  
8:00 236.3 F 11.1  
10:00 233.9 F 8.3  
12:00 238.1 F 10.3  
14:00 247.2 F 13.8  
16:00 260.1 F 20.9  
18:00 271.7 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.6 F 29.1  
2:00 272.5 F 29.1  
3:00 272.6 F 29.1  
4:00 272.6 F 29.2  
  
Min 270.8 F 28.1  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.0 F 0.0  
5:00 212.9 F 0.0  
10:00 226.3 F 0.0  
15:00 231.7 F 0.0  
20:00 229.3 F 0.0  
25:00 228.9 F 0.0  
30:00 227.1 F 0.0  
Drying Complete

FILLING: 2:45  
HEATING: 18:28  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:58:31  
CYCLE COMPLETE

09/27/2022 02:19 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_