
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000549
12:31 30/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:53
STERILIZING 3:59
136.0 C 220kPa 3:59
136.3 C 223kPa 4:29
136.4 C 225kPa 4:59
136.3 C 224kPa 5:29
136.3 C 223kPa 5:59
136.4 C 225kPa 6:29
136.4 C 224kPa 6:59
136.2 C 222kPa 7:29
136.2 C 222kPa 7:29
VENTING 7:29
AIR DRYING 8:24
CYCLE COMPLETE 8:28

Digital Signature #
147E9CB03F4C73A7
