
Midmark M11 - v1.0.5
Total Cycles: 524

Name: RITTER2
Sterilizer ID: V2313228
11/08/2021 02:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.6 F 0.0
2:00 177.3 F 3.6
4:00 199.4 F 9.3
6:00 217.9 F 7.1
8:00 233.5 F 11.5
10:00 237.3 F 10.2
12:00 237.2 F 9.7
14:00 252.4 F 17.0
16:00 264.5 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.4 F 29.4
2:00 272.7 F 29.9
3:00 271.7 F 29.3
4:00 272.5 F 29.5

Min 271.6 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 198.4 F 0.0
10:00 205.4 F 0.0
15:00 205.5 F 0.0
20:00 203.3 F 0.0
25:00 201.2 F 0.0
30:00 204.5 F 0.0

Drying Complete

FILLING: 2:11
HEATING: 17:49
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:10
CYCLE COMPLETE

11/08/2021 03:55 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____