
Midmark M11 - v1.0.5
Total Cycles: 1939

Name: RITTER2
Sterilizer ID: V2313228
01/17/2023 02:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.7 F 0.0
2:00 191.9 F 4.3
4:00 198.2 F 9.5
6:00 218.1 F 7.0
8:00 230.8 F 11.2
10:00 233.7 F 9.2
12:00 236.8 F 9.4
14:00 249.1 F 15.5
16:00 261.8 F 22.7
18:00 272.9 F 30.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.4 F 30.2
2:00 272.0 F 29.5
3:00 271.2 F 29.4
4:00 272.2 F 30.2

Min 271.1 F 29.2
Max 272.6 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 225.4 F 0.0
10:00 242.6 F 0.0
15:00 240.8 F 0.0
20:00 227.6 F 0.0
25:00 227.8 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 18:23
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:28
CYCLE COMPLETE

01/17/2023 03:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____