

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3769

Name: RITTER 1  
Sterilizer ID: V2178198  
09/14/2023 07:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.4 F 0.0  
2:00 70.6 F 0.0  
4:00 81.6 F 0.0  
6:00 111.9 F 0.5  
8:00 142.5 F 1.6  
10:00 169.6 F 2.8  
12:00 184.6 F 4.8  
14:00 206.6 F 8.2  
16:00 214.7 F 11.6  
18:00 219.9 F 4.3  
20:00 228.6 F 8.2  
22:00 237.2 F 12.2  
24:00 235.2 F 8.6  
26:00 244.8 F 12.3  
28:00 244.0 F 12.0  
30:00 255.4 F 17.9  
32:00 265.6 F 24.1  
34:00 271.3 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.3 F 28.6  
2:00 272.6 F 29.0  
3:00 271.8 F 28.3  
4:00 272.7 F 28.9  
  
Min 271.6 F 28.3  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.4 F 0.0  
5:00 212.7 F 0.0  
10:00 208.5 F 0.0  
15:00 202.2 F 0.0  
20:00 197.3 F 0.0  
25:00 192.8 F 0.0  
30:00 187.9 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 34:37  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:12:53  
CYCLE COMPLETE

09/14/2023 09:08 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_