
Midmark M11 - v1.0.0
Total Cycles: 3048

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2023 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.1 F 0.0
2:00 193.0 F 5.0
4:00 222.9 F 10.5
6:00 223.8 F 6.0
8:00 237.5 F 12.2
10:00 237.6 F 9.8
12:00 237.0 F 8.4
14:00 250.8 F 15.5
16:00 263.2 F 22.7
18:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.2 F 28.6
2:00 271.7 F 28.6
3:00 272.0 F 28.5
4:00 271.5 F 28.5

Min 271.1 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 215.2 F 0.0
10:00 237.9 F 0.0
15:00 234.4 F 0.0
20:00 225.6 F 0.0
25:00 222.3 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 2:22
HEATING: 18:03
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:57:41
CYCLE COMPLETE

03/28/2023 04:38 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____