

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1466

Name: RITTER 1  
Sterilizer ID: V2178198  
12/16/2021 11:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.3 F 0.0  
2:00 163.0 F 2.3  
4:00 182.5 F 5.9  
6:00 203.1 F 10.8  
8:00 222.2 F 5.2  
10:00 232.3 F 10.2  
12:00 232.4 F 7.8  
14:00 244.2 F 12.8  
16:00 241.7 F 11.2  
18:00 255.0 F 17.9  
20:00 266.4 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 272.8 F 29.0  
2:00 271.8 F 28.4  
3:00 271.5 F 28.7  
4:00 272.6 F 29.0

Min 271.0 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.6 F 0.0  
5:00 223.3 F 0.0  
10:00 200.3 F 0.0  
15:00 195.6 F 0.0  
20:00 186.8 F 0.0  
25:00 195.0 F 0.0  
Dry Cycle Aborted

FILLING: 1:10  
HEATING: 21:36  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 25:18  
TOTAL CYCLE: 00:55:13

12/16/2021 11:58 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02