

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4624

Name: RITTER 1  
Sterilizer ID: V2178198  
06/11/2024 08:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.3 F 0.0  
2:00 75.6 F 0.0  
4:00 85.3 F 0.1  
6:00 120.1 F 0.3  
8:00 148.0 F 1.1  
10:00 172.5 F 4.1  
12:00 188.9 F 7.3  
14:00 203.5 F 10.7  
16:00 213.0 F 7.5  
18:00 221.6 F 6.4  
20:00 231.2 F 10.2  
22:00 235.5 F 9.5  
24:00 241.0 F 11.4  
26:00 237.8 F 9.4  
28:00 251.0 F 15.7  
30:00 261.9 F 21.9  
32:00 272.2 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.5 F 28.7  
2:00 271.5 F 28.2  
3:00 272.6 F 28.8  
4:00 271.5 F 28.0  
  
Min 271.4 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.7 F 0.0  
5:00 206.0 F 0.0  
10:00 201.2 F 0.0  
15:00 204.8 F 0.0  
20:00 206.1 F 0.0  
25:00 208.0 F 0.0  
30:00 214.3 F 0.0  
  
Drying Complete

FILLING: 0:47  
HEATING: 32:31  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:10:45  
CYCLE COMPLETE

06/11/2024 09:16 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_