
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001736
15:03 15/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:46
STERILIZING 4:03
136.4 C 225kPa 4:03
136.3 C 223kPa 4:33
136.0 C 222kPa 5:03
136.4 C 226kPa 5:33
136.4 C 223kPa 6:03
136.4 C 225kPa 6:33
136.4 C 225kPa 7:03
136.4 C 226kPa 7:33
136.4 C 226kPa 7:33
VENTING 7:33
AIR DRYING 8:22
CYCLE COMPLETE 8:37

Digital Signature #
C843E0E66BB4893B
