
Midmark M11 - v1.0.0
Total Cycles: 2400

Name: RITTER 1
Sterilizer ID: V2178198
10/05/2022 07:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.5 F 0.0
2:00 81.5 F 0.0
4:00 86.2 F 0.0
6:00 115.0 F 0.8
8:00 139.5 F 2.3
10:00 166.6 F 4.6
12:00 186.8 F 7.7
14:00 204.3 F 11.2
16:00 216.6 F 5.0
18:00 223.9 F 8.3
20:00 235.5 F 12.4
22:00 234.6 F 8.7
24:00 242.0 F 11.4
26:00 243.9 F 12.2
28:00 255.3 F 18.1
30:00 265.5 F 24.4
32:00 272.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.5 F 29.0
2:00 272.0 F 28.9
3:00 272.4 F 28.8
4:00 272.1 F 29.1

Min 271.6 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.8 F 0.0
5:00 231.9 F 0.0
10:00 227.2 F 0.0
15:00 227.3 F 0.0
20:00 232.0 F 0.0
25:00 230.4 F 0.0
30:00 230.0 F 0.0

Drying Complete

FILLING: 0:41
HEATING: 32:31
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:10:27
CYCLE COMPLETE

10/05/2022 08:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____