

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5374

Name: RITTER 1  
Sterilizer ID: V2178198  
02/27/2025 11:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 103.4 F 0.0  
2:00 110.2 F 1.0  
4:00 132.1 F 3.3  
6:00 159.5 F 6.5  
8:00 179.2 F 10.4  
10:00 193.5 F 4.2  
12:00 210.5 F 5.9  
14:00 223.0 F 9.9  
16:00 228.9 F 9.1  
18:00 236.2 F 10.3  
20:00 235.0 F 8.4  
22:00 246.0 F 13.6  
24:00 257.8 F 20.0  
26:00 268.5 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.8 F 29.6  
2:00 272.1 F 28.9  
3:00 272.1 F 29.2  
4:00 272.8 F 29.4

Min 271.5 F 28.6  
Max 272.9 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.1 F 0.0  
5:00 186.3 F 0.0  
10:00 199.0 F 0.0  
15:00 205.6 F 0.0  
20:00 213.7 F 0.0  
25:00 216.3 F 0.0  
30:00 221.4 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 27:17  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:05:43  
CYCLE COMPLETE

02/27/2025 01:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_