
Midmark M11 - v1.0.0
Total Cycles: 5267

Name: RITTER 1
Sterilizer ID: V2178198
01/23/2025 12:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.4 F 0.0
2:00 166.7 F 3.0
4:00 185.7 F 7.1
6:00 202.4 F 11.6
8:00 218.2 F 3.9
10:00 226.1 F 8.2
12:00 238.7 F 12.8
14:00 237.5 F 9.4
16:00 239.0 F 9.7
18:00 246.9 F 13.5
20:00 258.2 F 19.7
22:00 268.8 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.2
1:00 272.3 F 28.9
2:00 271.4 F 28.2
3:00 271.7 F 28.8
4:00 272.5 F 29.1

Min 271.0 F 28.1
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 219.5 F 0.0
10:00 211.3 F 0.0
15:00 195.7 F 0.0
20:00 197.8 F 0.0
25:00 241.5 F 0.0
30:00 231.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:12
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:01:59
CYCLE COMPLETE

01/23/2025 01:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____