
Midmark M11 - v1.0.5
Total Cycles: 434

Name: RITTER2
Sterilizer ID: V2313228
10/08/2021 12:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.6 F 0.0
2:00 96.7 F 0.9
4:00 116.7 F 3.4
6:00 154.2 F 6.8
8:00 184.0 F 10.9
10:00 202.7 F 3.6
12:00 216.5 F 7.9
14:00 228.5 F 12.3
16:00 231.1 F 8.7
18:00 239.5 F 11.3
20:00 242.9 F 12.6
22:00 254.5 F 18.8
24:00 265.7 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 272.3 F 29.5
2:00 272.3 F 29.8
3:00 272.3 F 30.0
4:00 272.3 F 30.3

Min 271.7 F 29.5
Max 272.5 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.7 F 0.0
5:00 199.9 F 0.0
10:00 199.3 F 0.0
15:00 198.9 F 0.0
20:00 197.7 F 0.0
25:00 196.5 F 0.0
30:00 196.2 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 25:44
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:04:02
CYCLE COMPLETE

10/08/2021 01:24 PM
ACCEPTED: X
REJECTED:
Operator: U02