
Midmark M11 - v1.0.5
Total Cycles: 3234

Name: RITTER2
Sterilizer ID: V2313228
12/19/2023 03:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.7 F 0.0
2:00 103.6 F 0.3
4:00 126.3 F 1.9
6:00 157.3 F 4.8
8:00 179.8 F 8.4
10:00 199.2 F 12.6
12:00 212.6 F 5.9
14:00 225.8 F 10.1
16:00 230.0 F 8.9
18:00 238.2 F 11.6
20:00 237.2 F 9.9
22:00 250.1 F 16.0
24:00 261.3 F 22.4
26:00 271.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.5 F 29.7
2:00 271.2 F 29.2
3:00 272.6 F 29.8
4:00 271.3 F 29.1

Min 271.1 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 205.7 F 0.0
10:00 189.0 F 0.0
15:00 186.2 F 0.0
20:00 187.0 F 0.0
25:00 193.0 F 0.0
30:00 183.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:36
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:05:02
CYCLE COMPLETE

12/19/2023 04:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____