
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000858
11:15 23/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 2:57
STERILIZING 4:07
136.5 C 225kPa 4:07
136.4 C 225kPa 4:37
136.4 C 223kPa 5:07
136.4 C 225kPa 5:37
136.2 C 221kPa 6:07
136.4 C 224kPa 6:37
136.4 C 224kPa 7:07
136.2 C 222kPa 7:37
136.2 C 222kPa 7:37
VENTING 7:37
AIR DRYING 8:33
CYCLE COMPLETE 8:39

Digital Signature #
F1A03C94F7EF1157
