

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002452  
12:48 24/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:39  
PRESSURIZING 2:45  
STERILIZING 4:02  
136.4 C 224kPa 4:02  
136.4 C 225kPa 4:32  
136.3 C 222kPa 5:02  
136.1 C 223kPa 5:32  
136.4 C 225kPa 6:02  
136.4 C 222kPa 6:32  
136.4 C 224kPa 7:02  
136.4 C 225kPa 7:32  
136.4 C 225kPa 7:32  
VENTING 7:32  
AIR DRYING 8:19  
CYCLE COMPLETE 38:20  
-----

Digital Signature #  
7C1DA7A3D5633C7B  
-----