
Midmark M11 - v1.0.0
Total Cycles: 965

Name: RITTER 1
Sterilizer ID: V2178198
07/06/2021 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.3 F 0.0
2:00 160.7 F 1.0
4:00 172.3 F 3.5
6:00 188.3 F 7.1
8:00 203.5 F 11.8
10:00 216.5 F 5.2
12:00 229.3 F 9.9
14:00 237.5 F 9.2
16:00 241.1 F 11.8
18:00 238.7 F 9.9
20:00 251.5 F 16.2
22:00 262.5 F 22.5
24:00 272.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.2
2:00 271.9 F 28.5
3:00 271.8 F 28.9
4:00 272.4 F 29.1

Min 271.3 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.3 F 0.0
5:00 226.6 F 0.0
10:00 228.3 F 0.0
15:00 230.0 F 0.0
20:00 223.5 F 0.0
25:00 232.8 F 0.0
30:00 232.8 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:27
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:02:38
CYCLE COMPLETE

07/06/2021 09:13 AM
ACCEPTED: X
REJECTED:
Operator: U01