
Midmark M11 - v1.0.0
Total Cycles: 2000

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2022 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.6 F 0.0
2:00 164.0 F 2.9
4:00 167.6 F 6.6
6:00 194.3 F 11.1
8:00 214.5 F 5.0
10:00 228.6 F 10.1
12:00 233.2 F 7.6
14:00 240.7 F 12.0
16:00 238.3 F 10.0
18:00 251.9 F 16.6
20:00 263.3 F 23.3
22:00 272.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.5 F 29.5
2:00 271.8 F 28.8
3:00 271.3 F 28.9
4:00 272.4 F 29.5

Min 271.1 F 28.5
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 205.9 F 0.0
10:00 216.6 F 0.0
15:00 220.0 F 0.0
20:00 223.3 F 0.0
25:00 226.7 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:12
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:35
CYCLE COMPLETE

06/07/2022 11:19 AM
ACCEPTED: X
REJECTED:
Operator: U02