
Midmark M11 - v1.0.5
Total Cycles: 605

Name: RITTER2
Sterilizer ID: V2313228
12/06/2021 02:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.3 F 0.0
2:00 135.0 F 1.8
4:00 151.6 F 5.2
6:00 185.3 F 10.1
8:00 210.2 F 5.4
10:00 227.1 F 8.8
12:00 236.3 F 11.2
14:00 240.5 F 11.7
16:00 239.0 F 10.6
18:00 252.7 F 17.4
20:00 264.7 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.6 F 29.5
2:00 271.8 F 29.4
3:00 272.6 F 29.3
4:00 271.8 F 29.5

Min 271.4 F 28.8
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 240.6 F 0.0
10:00 218.8 F 0.0
15:00 209.1 F 0.0
20:00 209.1 F 0.0
25:00 220.8 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 21:49
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:28
CYCLE COMPLETE

12/06/2021 03:43 PM
ACCEPTED: X
REJECTED:
Operator: U02