

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004135  
11:14 03/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:25  
PRESSURIZING 2:32  
STERILIZING 3:27  
136.0 C 220kPa 3:27  
136.3 C 224kPa 3:57  
136.5 C 221kPa 4:27  
136.4 C 223kPa 4:57  
136.4 C 225kPa 5:27  
136.4 C 224kPa 5:57  
136.4 C 225kPa 6:27  
136.3 C 225kPa 6:57  
136.3 C 225kPa 6:57  
VENTING 6:57  
AIR DRYING 7:48  
CYCLE COMPLETE 8:01

-----  
Digital Signature #  
3F1B3E1094F3138C  
-----