
Midmark M11 - v1.0.0
Total Cycles: 3906

Name: RITTER 1
Sterilizer ID: V2178198
10/16/2023 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.4 F 0.0
2:00 187.1 F 3.2
4:00 195.8 F 7.9
6:00 219.2 F 12.1
8:00 229.8 F 8.6
10:00 240.7 F 10.9
12:00 241.0 F 11.8
14:00 239.7 F 10.4
16:00 253.5 F 17.3
18:00 265.2 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.4 F 29.0
2:00 271.5 F 28.6
3:00 272.4 F 29.1
4:00 271.5 F 28.4

Min 271.4 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 229.6 F 0.0
10:00 227.4 F 0.0
15:00 227.5 F 0.0
20:00 227.0 F 0.0
25:00 227.9 F 0.0
30:00 229.5 F 0.0

Drying Complete

FILLING: 1:44
HEATING: 19:46
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

10/16/2023 12:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____