
Midmark M11 - v1.0.0
Total Cycles: 3073

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2023 11:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.5 F 0.0
2:00 164.0 F 2.3
4:00 176.0 F 5.7
6:00 201.9 F 10.1
8:00 218.9 F 5.7
10:00 231.6 F 9.7
12:00 234.2 F 8.7
14:00 243.0 F 12.4
16:00 241.5 F 10.9
18:00 254.9 F 17.7
20:00 266.5 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.2 F 28.6
2:00 271.8 F 28.8
3:00 272.0 F 28.4
4:00 271.9 F 28.9

Min 271.3 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 223.6 F 0.0
10:00 230.0 F 0.0
15:00 229.7 F 0.0
20:00 220.9 F 0.0
25:00 217.0 F 0.0
30:00 215.9 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:39
CYCLE COMPLETE

04/04/2023 12:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____