
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001459
9:07 06/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:29
STERILIZING 4:21
136.0 C 220kPa 4:21
135.9 C 221kPa 4:51
136.3 C 223kPa 5:21
136.2 C 221kPa 5:51
136.4 C 223kPa 6:21
136.4 C 224kPa 6:51
136.4 C 224kPa 7:21
136.2 C 217kPa 7:51
136.2 C 217kPa 7:51
VENTING 7:52
AIR DRYING 8:47
CYCLE COMPLETE 8:54

Digital Signature #
186BA1760F172EB0
