
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001748
9:55 14/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:51
STERILIZING 3:50
136.4 C 220kPa 3:50
136.3 C 224kPa 4:20
136.2 C 221kPa 4:50
136.4 C 224kPa 5:20
136.4 C 224kPa 5:50
136.2 C 222kPa 6:20
136.4 C 224kPa 6:50
136.3 C 223kPa 7:20
136.3 C 223kPa 7:20
VENTING 7:20
AIR DRYING 8:09
CYCLE COMPLETE 8:32

Digital Signature #
21687E909034573C
