

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2901

Name: RITTER2  
Sterilizer ID: V2313228  
09/18/2023 03:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.8 F 0.0  
2:00 164.5 F 3.1  
4:00 186.8 F 7.8  
6:00 208.1 F 13.3  
8:00 223.1 F 8.1  
10:00 235.0 F 11.7  
12:00 239.9 F 12.1  
14:00 241.0 F 11.3  
16:00 254.6 F 18.3  
18:00 266.3 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.7 F 30.0  
2:00 272.0 F 29.3  
3:00 271.3 F 29.3  
4:00 272.7 F 30.0  
  
Min 271.2 F 28.9  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.2 F 0.0  
5:00 229.1 F 0.0  
10:00 215.3 F 0.0  
15:00 221.9 F 0.0  
20:00 177.9 F 0.0  
25:00 174.7 F 0.0  
30:00 172.5 F 0.0  
Drying Complete

FILLING: 1:50  
HEATING: 19:39  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 00:58:55  
CYCLE COMPLETE

09/18/2023 04:34 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_