

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1863

Name: RITTER 1  
Sterilizer ID: V2178198  
04/25/2022 12:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.0 F 0.0  
2:00 146.7 F 1.9  
4:00 168.0 F 5.1  
6:00 190.3 F 9.6  
8:00 216.6 F 7.0  
10:00 225.0 F 8.3  
12:00 235.1 F 13.3  
14:00 237.4 F 10.0  
16:00 236.7 F 8.3  
18:00 248.6 F 14.6  
20:00 260.7 F 21.3  
22:00 271.8 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.4 F 28.7  
2:00 272.3 F 29.0  
3:00 271.7 F 28.4  
4:00 272.7 F 29.1

Min 271.4 F 28.2  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.6 F 0.0  
5:00 212.1 F 0.0  
10:00 196.2 F 0.0  
15:00 199.7 F 0.0  
20:00 203.5 F 0.0  
25:00 226.3 F 0.0  
30:00 229.9 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 22:32  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:36  
CYCLE COMPLETE

04/25/2022 01:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_