
Midmark M11 - v1.0.5
Total Cycles: 2010

Name: RITTER2
Sterilizer ID: V2313228
02/06/2023 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.6 F 0.0
2:00 188.6 F 1.4
4:00 195.7 F 5.6
6:00 215.6 F 10.5
8:00 223.5 F 5.9
10:00 238.2 F 11.6
12:00 237.1 F 9.8
14:00 237.3 F 8.8
16:00 251.0 F 16.0
18:00 263.1 F 23.1
20:00 271.9 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.5 F 29.3
2:00 271.4 F 28.5
3:00 272.3 F 29.5
4:00 272.0 F 29.0

Min 271.3 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 230.6 F 0.0
10:00 241.9 F 0.0
15:00 229.9 F 0.0
20:00 227.7 F 0.0
25:00 230.7 F 0.0
30:00 231.8 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 20:09
STERILIZING: 4:00
VENTING: 2:60
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

02/06/2023 12:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____