
Midmark M11 - v1.0.0
Total Cycles: 2822

Name: RITTER 1
Sterilizer ID: V2178198
01/30/2023 08:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.8 F 0.0
2:00 76.4 F 0.0
4:00 82.7 F 0.1
6:00 111.6 F 1.9
8:00 147.4 F 4.5
10:00 171.1 F 7.7
12:00 201.2 F 11.0
14:00 211.3 F 3.8
16:00 216.0 F 7.4
18:00 227.1 F 11.3
20:00 228.2 F 6.5
22:00 240.7 F 11.8
24:00 237.5 F 9.3
26:00 250.4 F 15.4
28:00 261.2 F 21.6
30:00 271.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 271.9 F 28.4
2:00 272.4 F 28.8
3:00 272.1 F 29.0
4:00 271.6 F 28.3

Min 271.4 F 28.2
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 216.8 F 0.0
10:00 213.5 F 0.0
15:00 211.7 F 0.0
20:00 220.8 F 0.0
25:00 226.4 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:34
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:36
CYCLE COMPLETE

01/30/2023 09:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____