
Midmark M11 - v1.0.5
Total Cycles: 1861

Name: RITTER2
Sterilizer ID: V2313228
12/20/2022 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.7 F 0.0
2:00 162.9 F 0.0
4:00 181.0 F 1.1
6:00 204.6 F 6.1
8:00 216.9 F 11.4
10:00 226.8 F 7.4
12:00 240.4 F 13.4
14:00 241.2 F 11.8
16:00 240.9 F 11.0
18:00 255.0 F 18.2
20:00 266.9 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.6 F 29.6
2:00 272.0 F 28.9
3:00 271.6 F 29.0
4:00 272.6 F 29.6

Min 271.3 F 28.6
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.1 F 0.0
5:00 221.2 F 0.0
10:00 235.2 F 0.0
15:00 237.3 F 0.0
20:00 225.6 F 0.0
25:00 223.9 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 21:30
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

12/20/2022 11:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____