
Midmark M11 - v1.0.0
Total Cycles: 4751

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2024 02:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.4 F 0.0
2:00 99.4 F 0.8
4:00 126.2 F 2.9
6:00 157.8 F 5.7
8:00 193.9 F 9.4
10:00 211.8 F 11.2
12:00 219.4 F 5.3
14:00 228.4 F 9.0
16:00 236.6 F 13.1
18:00 236.3 F 9.0
20:00 241.8 F 11.0
22:00 244.0 F 12.1
24:00 255.3 F 17.9
26:00 265.6 F 24.1
28:00 271.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 271.9 F 28.4
2:00 272.3 F 28.7
3:00 272.7 F 29.0
4:00 271.8 F 29.0

Min 271.3 F 28.4
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 246.7 F 0.0
10:00 196.8 F 0.0
15:00 219.4 F 0.0
20:00 229.5 F 0.0
25:00 220.6 F 0.0
30:00 218.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 28:30
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:07:09
CYCLE COMPLETE

07/19/2024 03:29 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____