
Midmark M11 - v1.0.5
Total Cycles: 3587

Name: RITTER2
Sterilizer ID: V2313228
05/03/2024 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.2 F 0.0
2:00 142.9 F 1.6
4:00 154.5 F 4.2
6:00 176.4 F 7.5
8:00 197.5 F 11.7
10:00 223.9 F 5.5
12:00 230.4 F 9.8
14:00 235.5 F 9.6
16:00 240.3 F 11.3
18:00 237.3 F 9.4
20:00 250.3 F 15.6
22:00 261.1 F 21.7
24:00 270.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.9 F 29.7
2:00 271.6 F 29.6
3:00 272.5 F 29.6
4:00 271.6 F 29.6

Min 271.0 F 29.2
Max 273.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 234.2 F 0.0
10:00 193.5 F 0.0
15:00 181.4 F 0.0
20:00 177.2 F 0.0
25:00 176.1 F 0.0
30:00 175.2 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 25:47
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:04:33
CYCLE COMPLETE

05/03/2024 11:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____