
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000206
14:57 21/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:22
STERILIZING 4:45
136.0 C 220kPa 4:45
136.4 C 232kPa 5:15
136.2 C 232kPa 5:45
136.0 C 231kPa 6:15
136.4 C 233kPa 6:45
136.4 C 233kPa 7:15
136.3 C 232kPa 7:45
136.3 C 231kPa 8:15
136.3 C 231kPa 8:15
VENTING 8:15
AIR DRYING 9:19
CYCLE COMPLETE 24:19

Digital Signature #
354997097F94905E
