
Midmark M11 - v1.0.5
Total Cycles: 4212

Name: RITTER2
Sterilizer ID: V2313228
12/13/2024 01:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.6 F 0.0
2:00 140.3 F 1.5
4:00 155.0 F 4.2
6:00 178.2 F 8.0
8:00 197.6 F 12.6
10:00 217.7 F 6.2
12:00 234.2 F 10.9
14:00 230.8 F 7.1
16:00 242.2 F 12.9
18:00 241.8 F 11.4
20:00 253.9 F 17.5
22:00 264.5 F 24.0
24:00 270.8 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 272.6 F 29.3
1:00 272.7 F 29.5
2:00 272.9 F 29.7
3:00 272.5 F 29.9
4:00 272.0 F 29.3

Min 271.8 F 29.2
Max 273.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.3 F 0.0
5:00 235.2 F 0.0
10:00 227.5 F 0.0
15:00 182.8 F 0.0
20:00 180.1 F 0.0
25:00 178.5 F 0.0
30:00 178.0 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 24:32
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:03:19
CYCLE COMPLETE

12/13/2024 02:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____