
Midmark M11 - v1.0.0
Total Cycles: 1791

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2022 02:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.1 F 0.0
2:00 123.5 F 0.3
4:00 151.9 F 3.2
6:00 182.1 F 7.0
8:00 198.7 F 11.5
10:00 215.3 F 4.3
12:00 228.1 F 8.7
14:00 238.8 F 13.4
16:00 238.0 F 10.1
18:00 236.3 F 8.4
20:00 248.3 F 14.4
22:00 260.2 F 21.0
24:00 270.8 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.7 F 29.0
2:00 271.6 F 28.2
3:00 272.6 F 29.2
4:00 272.1 F 28.6

Min 271.3 F 28.1
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 221.8 F 0.0
10:00 220.3 F 0.0
15:00 216.3 F 0.0
20:00 213.8 F 0.0
25:00 225.7 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:46
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:02:47
CYCLE COMPLETE

04/04/2022 03:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____