
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006074
14:32 03/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:28
STERILIZING 3:19
136.0 C 220kPa 3:19
136.2 C 222kPa 3:49
136.3 C 223kPa 4:19
136.2 C 222kPa 4:49
136.3 C 221kPa 5:19
136.3 C 225kPa 5:49
136.3 C 223kPa 6:19
136.4 C 222kPa 6:49
136.4 C 222kPa 6:49
VENTING 6:49
AIR DRYING 7:43
CYCLE COMPLETE 9:45

Digital Signature #
3A04E11FBBC1DA0C
