
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006081
9:27 09/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:55
STERILIZING 3:59
136.0 C 220kPa 3:59
136.0 C 222kPa 4:29
136.3 C 220kPa 4:59
136.3 C 223kPa 5:29
136.4 C 222kPa 5:59
135.9 C 221kPa 6:29
136.4 C 224kPa 6:59
136.3 C 221kPa 7:29
136.3 C 221kPa 7:29
VENTING 7:29
AIR DRYING 8:20
CYCLE COMPLETE 21:31

Digital Signature #
58CF5EC3F77DCACB
