

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2268

Name: RITTER 1  
Sterilizer ID: V2178198  
08/30/2022 02:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.3 F 0.0  
2:00 137.3 F 0.4  
4:00 171.6 F 2.8  
6:00 195.1 F 7.3  
8:00 210.4 F 12.7  
10:00 222.6 F 7.7  
12:00 236.0 F 13.0  
14:00 236.9 F 10.3  
16:00 234.9 F 8.3  
18:00 249.9 F 15.6  
20:00 262.1 F 22.6  
22:00 272.6 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.6 F 29.3  
2:00 271.6 F 28.7  
3:00 272.7 F 29.4  
4:00 271.8 F 28.6

Min 271.6 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.2 F 0.0  
5:00 201.0 F 0.0  
10:00 210.2 F 0.0  
15:00 213.3 F 0.0  
20:00 217.2 F 0.0  
25:00 221.8 F 0.0  
30:00 223.2 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 22:16  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:01:05  
CYCLE COMPLETE

08/30/2022 03:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_