
Midmark M11 - v1.0.0
Total Cycles: 3454

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2023 02:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.9 F 0.0
2:00 127.9 F 0.3
4:00 150.8 F 2.0
6:00 174.5 F 5.2
8:00 195.1 F 8.9
10:00 210.4 F 13.4
12:00 219.8 F 6.3
14:00 230.8 F 10.5
16:00 232.2 F 8.1
18:00 240.8 F 11.8
20:00 238.3 F 9.8
22:00 250.7 F 15.7
24:00 261.4 F 21.8
26:00 271.3 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.1
2:00 271.7 F 28.4
3:00 272.4 F 29.2
4:00 272.4 F 28.9

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 215.8 F 0.0
10:00 219.8 F 0.0
15:00 220.6 F 0.0
20:00 220.3 F 0.0
25:00 218.6 F 0.0
30:00 219.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:41
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:05:07
CYCLE COMPLETE

06/27/2023 03:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____